

Certified Calibration Fluid

Batch No. 6982121

Part No. CALI0000002

Exp. Date: See bottle

100 cP

Grace Instrument Company

Houston, TX 77055

(713) 783-1560

| [°] F/ [°] C | Centipoise | [°] F/ [°] C | Centipoise | [°] F/ [°] C | Centipoise | [°] F/ [°] C | Centipoise |
|--------------------------------|------------|--------------------------------|------------|--------------------------------|------------|--------------------------------|------------|
| 68 / 20 | 104.70 | 77 / 25 | 94.26 | 86 / 30 | 85.28 | 95 / 35 | 77.48 |
| 20.1 | 104.49 | 25.1 | 94.08 | 30.1 | 85.12 | 35.1 | 77.34 |
| 20.2 | 104.28 | 25.2 | 93.90 | 30.2 | 84.97 | 35.2 | 77.20 |
| 20.3 | 104.07 | 25.3 | 93.72 | 30.3 | 84.81 | 35.3 | 77.07 |
| 20.4 | 103.86 | 25.4 | 93.54 | 30.4 | 84.66 | 35.4 | 76.93 |
| 20.5 | 103.66 | 25.5 | 93.36 | 30.5 | 84.50 | 35.5 | 76.79 |
| 69 / 20.6 | 103.45 | 78 / 25.6 | 93.18 | 87 / 30.6 | 84.34 | 96 / 35.6 | 76.65 |
| 20.7 | 103.24 | 25.7 | 93.00 | 30.7 | 84.19 | 35.7 | 76.51 |
| 20.8 | 103.03 | 25.8 | 92.82 | 30.8 | 84.03 | 35.8 | 76.38 |
| 20.9 | 102.82 | 25.9 | 92.64 | 30.9 | 83.88 | 35.9 | 76.24 |
| 21.0 | 102.61 | 26.0 | 92.46 | 31.0 | 83.72 | 36.0 | 76.10 |
| 70 / 21.1 | 102.40 | 79 / 26.1 | 92.28 | 88 / 31.1 | 83.56 | 97 / 36.1 | 75.96 |
| 21.2 | 102.19 | 26.2 | 92.10 | 31.2 | 83.41 | 36.2 | 75.82 |
| 21.3 | 101.99 | 26.3 | 91.93 | 31.3 | 83.25 | 36.3 | 75.69 |
| 21.4 | 101.78 | 26.4 | 91.75 | 31.4 | 83.10 | 36.4 | 75.55 |
| 21.5 | 101.57 | 26.5 | 91.57 | 31.5 | 82.94 | 36.5 | 75.41 |
| 21.6 | 101.36 | 26.6 | 91.39 | 31.6 | 82.78 | 36.6 | 75.27 |
| 71 / 21.7 | 101.15 | 80 / 26.7 | 91.21 | 89 / 31.7 | 82.63 | 98 / 36.7 | 75.13 |
| 21.8 | 100.94 | 26.8 | 91.03 | 31.8 | 82.47 | 36.8 | 75.00 |
| 21.9 | 100.73 | 26.9 | 90.85 | 31.9 | 82.32 | 36.9 | 74.86 |
| 22.0 | 100.52 | 27.0 | 90.67 | 32.0 | 82.16 | 37.0 | 74.72 |
| 22.1 | 100.32 | 27.1 | 90.49 | 32.1 | 82.00 | 37.1 | 74.58 |
| 72 / 22.2 | 100.11 | 81 / 27.2 | 90.31 | 90 / 32.2 | 81.85 | 99 / 37.2 | 74.44 |
| 22.3 | 99.90 | 27.3 | 90.13 | 32.3 | 81.69 | 37.3 | 74.31 |
| 22.4 | 99.69 | 27.4 | 89.95 | 32.4 | 81.54 | 37.4 | 74.17 |
| 22.5 | 99.48 | 27.5 | 89.77 | 32.5 | 81.38 | 37.5 | 74.03 |
| 22.6 | 99.27 | 27.6 | 89.59 | 32.6 | 81.22 | 37.6 | 73.89 |
| 22.7 | 99.06 | 27.7 | 89.41 | 32.7 | 81.07 | 37.7 | 73.75 |
| 73 / 22.8 | 98.85 | 82 / 27.8 | 89.23 | 91 / 32.8 | 80.91 | 100 / 37.8 | 73.62 |
| 22.9 | 98.64 | 27.9 | 89.05 | 32.9 | 80.76 | 37.9 | 73.48 |
| 23.0 | 98.44 | 28.0 | 88.87 | 33.0 | 80.60 | 38.0 | 73.34 |
| 23.1 | 98.23 | 28.1 | 88.69 | 33.1 | 80.44 | 38.1 | 73.20 |
| 23.2 | 98.02 | 28.2 | 88.51 | 33.2 | 80.29 | 38.2 | 73.06 |
| 74 / 23.3 | 97.81 | 83 / 28.3 | 88.33 | 92 / 33.3 | 80.13 | 101 / 38.3 | 72.93 |
| 23.4 | 97.60 | 28.4 | 88.15 | 33.4 | 79.98 | 38.4 | 72.79 |
| 23.5 | 97.39 | 28.5 | 87.97 | 33.5 | 79.82 | 38.5 | 72.65 |
| 23.6 | 97.18 | 28.6 | 87.79 | 33.6 | 79.66 | 38.6 | 72.51 |
| 23.7 | 96.97 | 28.7 | 87.61 | 33.7 | 79.51 | 38.7 | 72.37 |
| 23.8 | 96.77 | 28.8 | 87.44 | 33.8 | 79.35 | 38.8 | 72.24 |
| 75 / 23.9 | 96.56 | 84 / 28.9 | 87.26 | 93 / 33.9 | 79.20 | 102 / 38.9 | 72.10 |
| 24.0 | 96.35 | 29.0 | 87.08 | 34.0 | 79.04 | 39.0 | 71.96 |
| 24.1 | 96.14 | 29.1 | 86.90 | 34.1 | 78.88 | 39.1 | 71.82 |
| 24.2 | 95.93 | 29.2 | 86.72 | 34.2 | 78.73 | 39.2 | 71.68 |
| 24.3 | 95.72 | 29.3 | 86.54 | 34.3 | 78.57 | 39.3 | 71.55 |
| 76 / 24.4 | 95.51 | 85 / 29.4 | 86.36 | 94 / 34.4 | 78.42 | 103 / 39.4 | 71.41 |
| 24.5 | 95.30 | 29.5 | 86.18 | 34.5 | 78.26 | 39.5 | 71.27 |
| 24.6 | 95.10 | 29.6 | 86.00 | 34.6 | 78.10 | 39.6 | 71.13 |
| 24.7 | 94.89 | 29.7 | 85.82 | 34.7 | 77.95 | 39.7 | 70.99 |
| 24.8 | 94.68 | 29.8 | 85.64 | 34.8 | 77.79 | 39.8 | 70.86 |
| 24.9 | 94.47 | 29.9 | 85.46 | 34.9 | 77.64 | 39.9 | 70.72 |
| | | | | | | 104 / 40 | 70.58 |