

Certified Calibration Fluid
 Batch No. 6481973
 Part No. CALI0000001

Grace Instrument Company
 Exp. Date: See bottle Houston, TX 77055
 50 cP (713) 783-1560

°F/°C	Centipoise	°F/°C	Centipoise	°F/°C	Centipoise	°F/°C	Centipoise
68 / 20	53.01	77 / 25	47.85	86 / 30	43.30	95 / 35	39.33
20.1	52.91	25.1	47.76	30.1	43.22	35.1	39.26
20.2	52.80	25.2	47.67	30.2	43.14	35.2	39.19
20.3	52.70	25.3	47.58	30.3	43.06	35.3	39.12
20.4	52.60	25.4	47.49	30.4	42.98	35.4	39.05
20.5	52.49	25.5	47.40	30.5	42.90	35.5	38.98
69 / 20.6	52.39	78 / 25.6	47.30	87 / 30.6	42.82	96 / 35.6	38.91
20.7	52.29	25.7	47.21	30.7	42.74	35.7	38.84
20.8	52.18	25.8	47.12	30.8	42.66	35.8	38.77
20.9	52.08	25.9	47.03	30.9	42.59	35.9	38.70
21.0	51.98	26.0	46.94	31.0	42.51	36.0	38.63
70 / 21.1	51.87	79 / 26.1	46.85	88 / 31.1	42.43	97 / 36.1	38.56
21.2	51.77	26.2	46.76	31.2	42.35	36.2	38.49
21.3	51.67	26.3	46.67	31.3	42.27	36.3	38.42
21.4	51.57	26.4	46.58	31.4	42.19	36.4	38.35
21.5	51.46	26.5	46.49	31.5	42.11	36.5	38.28
21.6	51.36	26.6	46.39	31.6	42.03	36.6	38.21
71 / 21.7	51.26	80 / 26.7	46.30	89 / 31.7	41.95	98 / 36.7	38.14
21.8	51.15	26.8	46.21	31.8	41.87	36.8	38.07
21.9	51.05	26.9	46.12	31.9	41.79	36.9	38.00
22.0	50.95	27.0	46.03	32.0	41.71	37.0	37.93
22.1	50.84	27.1	45.94	32.1	41.63	37.1	37.86
72 / 22.2	50.74	81 / 27.2	45.85	90 / 32.2	41.55	99 / 37.2	37.79
22.3	50.64	27.3	45.76	32.3	41.47	37.3	37.72
22.4	50.53	27.4	45.67	32.4	41.39	37.4	37.65
22.5	50.43	27.5	45.58	32.5	41.32	37.5	37.58
22.6	50.33	27.6	45.48	32.6	41.24	37.6	37.50
22.7	50.22	27.7	45.39	32.7	41.16	37.7	37.43
73 / 22.8	50.12	82 / 27.8	45.30	91 / 32.8	41.08	100 / 37.8	37.36
22.9	50.02	27.9	45.21	32.9	41.00	37.9	37.29
23.0	49.91	28.0	45.12	33.0	40.92	38.0	37.22
23.1	49.81	28.1	45.03	33.1	40.84	38.1	37.15
23.2	49.71	28.2	44.94	33.2	40.76	38.2	37.08
74 / 23.3	49.60	83 / 28.3	44.85	92 / 33.3	40.68	101 / 38.3	37.01
23.4	49.50	28.4	44.76	33.4	40.60	38.4	36.94
23.5	49.40	28.5	44.67	33.5	40.52	38.5	36.87
23.6	49.29	28.6	44.57	33.6	40.44	38.6	36.80
23.7	49.19	28.7	44.48	33.7	40.36	38.7	36.73
23.8	49.09	28.8	44.39	33.8	40.28	38.8	36.66
75 / 23.9	48.99	84 / 28.9	44.30	93 / 33.9	40.20	102 / 38.9	36.59
24.0	48.88	29.0	44.21	34.0	40.12	39.0	36.52
24.1	48.78	29.1	44.12	34.1	40.04	39.1	36.45
24.2	48.68	29.2	44.03	34.2	39.97	39.2	36.38
24.3	48.57	29.3	43.94	34.3	39.89	39.3	36.31
76 / 24.4	48.47	85 / 29.4	43.85	94 / 34.4	39.81	103 / 39.4	36.24
24.5	48.37	29.5	43.76	34.5	39.73	39.5	36.17
24.6	48.26	29.6	43.66	34.6	39.65	39.6	36.10
24.7	48.16	29.7	43.57	34.7	39.57	39.7	36.03
24.8	48.06	29.8	43.48	34.8	39.49	39.8	35.96
24.9	47.95	29.9	43.39	34.9	39.41	39.9	35.89
						104 / 40	35.82